
Generator Constant potential, microprocessor-controlled  Working frequency 145-230 KHz (typically 175 KHz)  Focal spot 0.4 mm (IEC 336)  Anode current 4/8 mA  Voltage at X-ray tube 60 / 65 / 70 kV (*)  Exposure time 0.020 – 1.000 seconds, R*10 and R*20 scale  Source-skin distance 20 and 30 cm  Irradiated field 35 x 45 mm (with rectangular cone for size 2 sensors), Ø 60 mm and Ø 55 mm (with round cone)  Additional collimators 31 x 41 mm and 22 x 35 mm, for size 1 and size 0 sensors  Total filtration 2 mm @ 60 kV / 2 mm @ 75 kV / 2 mm @ 70 kV (*)  Power supply 50/60 Hz, 115-120 V AC ±10% or 230-240 V AC ±10%  Duty Cycle Continuous operation with self-adjustment up to 1s/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iRYS (compliant with ISDP©10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS  ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)		TECHNICAL DATA
Working frequency 145-230 KHz (typically 175 KHz)  Focal spot 0.4 mm (IEC 336)  Anode current 4 / 8 mA  Voltage at X-ray tube 60 / 65 / 70 kV (*)  Exposure time 0.020 – 1.000 seconds, R*10 and R*20 scale  Source-skin distance 20 and 30 cm  Irradiated field 35 x 45 mm (with rectangular cone for size 2 sensors), Ø 60 mm and Ø 55 mm (with round cone)  Additional collimators 31 x 41 mm and 22 x 35 mm, for size 1 and size 0 sensors  Total filtration 2 mm @ 00 kV / 2 mm @ 65 kV / 2 mm @ 70 kV (*)  Power supply 50 /60 Hz, 115-120 V AC ±10% or 230-240 V AC ±10%  Duty Cycle Continuous operation with self-adjustment up to 1 x/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Vewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) (Capture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) (Septure for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) (Septure for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) (Septure for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) (Septure for automatic saving of RX DC connections with EN ISO/IEC17065:2012 - certificate number 201900310 (RYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft* Windows* 10.11 Professional 64 bit  Frocessor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  Gaphics card		TECHNICAL DATA
Focal spot 0.4 mm (IEC 336) Anode current 4 /8 mA Voltage at X-ray tube 60 / 65 / 70 kV (*) Exposure time 0.020 – 1.000 seconds, R*10 and R*20 scale Source-skin distance 20 and 30 cm Irradiated field 35 x 45 mm (with rectangular cone for size 2 sensors), Ø 60 mm and Ø 55 mm (with round cone) Additional collimators 31 x 41 mm and 22 x 35 mm, for size 1 and size 0 sensors Total filtration 2 mm @ 60 kV / 2 mm @ 65 kV / 2 mm @ 70 kV (*) Power supply 50/60 Hz, 115-120 V AC ± 10% or 230-240 V AC ± 10% Duty Cycle Continuous operation with self-adjustment up to 1s/80s total Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology) Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm Maximum arm extension 230 cm, from wall Dose delivered Vewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC Image management software (for PC) iRYS (compliant with ISDP010003-2020 in accordance with EN ISO/IEC17065-2012 - certificate number 201900310 (Price)  Protocols supported in iRYS ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft* Windows* 10.11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  Graphics card		
Anode current 4 / 8 mA  Voltage at X-ray tube 60 / 65 / 70 kV (*)  Exposure time 0.020 – 1.000 seconds, R*10 and R*20 scale  Source-skin distance 20 and 30 cm  Irradiated field 35 x 45 mm (with rectangular cone for size 2 sensors), Ø 60 mm and Ø 55 mm (with round cone)  Additional collimators 31 x 41 mm and 22 x 35 mm, for size 1 and size 0 sensors  Total filtration 2 mm @ 60 kV / 2 mm @ 76 kV / 7  Power supply 50 / 60 Hz, 115-120 V AC ±10% or 230-240 V AC ±10%  Duty Cycle Continuous o peration with self-adjustment up to 1s/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iRYS (compliant with ISDP®10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS (CoM 3.0, TWAIN), VDDS  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList, MPPS, Query/Retrieve)  X-ray log iRYS - feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft* Windows* 10.11 Professional 64 bit  Frocessor Intel Core 13 or higher  Hard Disk 100 GB SSD (250 GB recommended)  Graphics card		
Voltage at X-ray tube 60 / 65 / 70 kV (*)  Exposure time 0.020 – 1.000 seconds, R10 and R20 scale  Source-skin distance 20 and 30 cm  Irradiated field 35 x 45 mm (with rectangular cone for size 2 sensors), Ø 60 mm and Ø 55 mm (with round cone)  Additional collimators 31 x 41 mm and 22 x 35 mm, for size 1 and size 0 sensors  Total filtration 2 mm @ 60 kV / 2 mm @ 65 kV / 2 mm @ 70 kV (*)  Power supply 50 /60 Hz, 115-120 V AC ± 10% or 230-240 V AC ± 10%  Duty Cycle Continuous operation with self-adjustment up to 1s/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Protocols supported in iRYS (COM 3.0, TWAIN, VDDS)  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft* Windows* 10.11 Professional 64 bit  Intel Core i3 or higher  Hard Disk 106 G8 B8 recommended)  Gaphics card Discrete 3D Video Card or integrated GPU	· · · · · · · · · · · · · · · · · · ·	V - 11114
Exposure time 0.020 – 1.000 seconds, R10 and R20 scale  Source-skin distance 20 and 30 cm  Irradiated field 35 x 45 mm (with rectangular cone for size 2 sensors), Ø 60 mm and Ø 55 mm (with round cone)  Additional collimators 31 x 41 mm and 22 x 35 mm, for size 1 and size 0 sensors  Total filtration 2 mm @ 60 kW / 2 mm @ 70 kW (*)  Power supply 50/60 Hz, 115-120 V AC ±10% or 230-240 V AC ±10%  Duty Cycle Continuous operation with self-adjustment up to 1s/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm ~ 60 cm ~ 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iRYS (compliant with ISDP®10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS (CM) 3.0, TWAIN, VDDS  DICOM Node Connectivity iRYS -IHE compliant (Pint; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft* Windows* 10.11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 66 Be recommended)  4 46 (8 GB recommended)  Discrete 3D Video Card or integrated GPU		4/8 mA
Source-skin distance 20 and 30 cm  Irradiated field 35 x 45 mm (with rectangular cone for size 2 sensors), Ø 60 mm and Ø 55 mm (with round cone)  Additional collimators 31 x 41 mm and 22 x 35 mm, for size 1 and size 0 sensors  Total filtration 2 mm @ 60 kV / 2 mm @ 65 kV / 2 mm @ 70 kV (*)  Power supply 50/60 Hz, 115-120 V AC ±10% or 230-240 V AC ±10%  Duty Cycle Continuous operation with self-adjustment up to 1s/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Vewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iProtocols supported in iRYS (compliant with ISDP©1003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310; iPad iRYS viewer App (free)  Protocols supported in iRYS ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList, MPPS; Query/Retrieve)  X-ray log iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft® Windows* 10.11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  Graphics card Discrete 3D Video Card or integrated GPU	Voltage at X-ray tube	60 / 65 / 70 kV (*)
Irradiated field 35 x 45 mm (with rectangular cone for size 2 sensors), Ø 60 mm and Ø 55 mm (with round cone)  Additional collimators 31 x 41 mm and 22 x 35 mm, for size 1 and size 0 sensors  Total filtration 2 mm @ 60 kV / 2 mm @ 65 kV / 2 mm @ 70 kV (*)  Power supply 50/60 Hz, 115-120 V AC ±1096 or 230-240 V AC ±1096  Duty Cycle Continuous operation with self-adjustment up to 1s/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iPy iPy selver App (free)  Protocols supported in iRYS (compliant with ISDP® 10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft® Windows® 10.11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  RAM 4 GB (8 GB recommended)  Discrete 3D Video Card or integrated GPU	Exposure time	0.020 – 1.000 seconds, R'10 and R'20 scale
Additional collimators 31 x 41 mm and 22 x 35 mm, for size 1 and size 0 sensors  Total filtration 2 mm @ 60 kV / 2 mm @ 65 kV / 2 mm @ 70 kV (*)  Power supply 50/60 Hz, 115-120 V AC ±10% or 230-240 V AC ±10%  Duty Cycle Continuous operation with self-adjustment up to 1s/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) iRYS (compliant with ISDP® 10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  Withilmum System Requisites  Supported operating systems Microsoft* Windows* 10.11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  RAM 4 GB (8 GB recommended)  Graphics card	Source-skin distance	20 and 30 cm
Total filtration 2 mm @ 60 kV / 2 mm @ 65 kV / 2 mm @ 70 kV (*)  Power supply 50/60 Hz, 115-120 V AC ±10% or 230-240 V AC ±10%  Duty Cycle Continuous operation with self-adjustment up to 1s/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iRYS (compliant with ISDP®10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or Compliant With ISDP®1011 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  Graphics card Discrete 3D Video Card or integrated GPU	Irradiated field	$35 \times 45 \text{ mm}$ (with rectangular cone for size 2 sensors), Ø 60 mm and Ø 55 mm (with round cone)
Power supply 50/60 Hz, 115-120 V AC ±10% or 230-240 V AC ±10%  Duty Cycle Continuous operation with self-adjustment up to 1s/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iRYS (compliant with ISDP©10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS iCOM 3.0, TWAIN, VDDS  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or Compliant Minions)  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft* Windows* 10.11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  RAM 4 GB (8 GB recommended)  Discrete 3D Video Card or integrated GPU	Additional collimators	31 x 41 mm and 22 x 35 mm, for size 1 and size 0 sensors
Duty Cycle Continuous operation with self-adjustment up to 1s/80s total  Stability Automatic lock/release, with touch-sensitive activation (HyperSphere technology)  Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) ICapture for automatic saving of RX DC exposure parameters on PC Image management software (for PC) IRYS (compliant with ISDP@10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 (iPad iRYS viewer App (free)  Protocols supported in iRYS ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity IRYS -IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log IRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft® Windows® 10.11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  Graphics card Discrete 3D Video Card or integrated GPU	Total filtration	2 mm @ 60 kV / 2 mm @ 65 kV / 2 mm @ 70 kV (*)
Stability       Automatic lock/release, with touch-sensitive activation (HyperSphere technology)         Arms       Available in 3 lengths: 40 cm - 60 cm - 90 cm         Maximum arm extension       230 cm, from wall         Dose delivered       Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory         PC connection cable       Serial with USB adapter available in various lengths         SOFTWARE         Acquisition software (for PC)       iCapture for automatic saving of RX DC exposure parameters on PC         Image management software (for PC)       iRYS (compliant with ISDP@10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)         Protocols supported in iRYS       ICOM 3.0, TWAIN, VDDS         DICOM Node Connectivity       iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)         X-ray log       iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C         MINIMUM SYSTEM REQUISITES         Supported operating systems       Microsoft* Windows* 10.11 Professional 64 bit         Processor       Intel Core i3 or higher         Hard Disk       100 GB SSD (250 GB recommended)         RAM       4 GB (8 GB recommended)         Graphics card       Discrete 3D Video Card or integra	Power supply	50/60 Hz, 115-120 V AC ±10% or 230-240 V AC ±10%
Arms Available in 3 lengths: 40 cm - 60 cm - 90 cm  Maximum arm extension 230 cm, from wall  Dose delivered Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iRYS (compliant with ISDP©10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft* Windows* 10.11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  RAM 4 GB (8 GB recommended)  Graphics card Discrete 3D Video Card or integrated GPU	Duty Cycle	Continuous operation with self-adjustment up to 1s/80s total
Maximum arm extension  230 cm, from wall  Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via "RX DC connect" (optional) accessory  PC connection cable  Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC)  ICapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC)  IRYS (compliant with ISDP®10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS  ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity  IRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log  iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems  Microsoft® Windows® 10.11 Professional 64 bit  Processor  Intel Core i3 or higher  Hard Disk  100 GB SSD (250 GB recommended)  RAM  4 GB (8 GB recommended)  Discrete 3D Video Card or integrated GPU	Stability	Automatic lock/release, with touch-sensitive activation (HyperSphere technology)
Dose delivered  Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via 'RX DC connect' (optional) accessory  PC connection cable  Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC)  iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC)  iRYS (compliant with ISDP®10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS  ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity  iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log  iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant of the X-ray images of each examination (exportable in PDF or Compliant	Arms	Available in 3 lengths: 40 cm - 60 cm - 90 cm
#RX DC connect" (optional) accessory  PC connection cable  Serial with USB adapter available in various lengths  SOFTWARE  Acquisition software (for PC)  Image management sof	Maximum arm extension	230 cm, from wall
SOFTWARE  Acquisition software (for PC) iCapture for automatic saving of RX DC exposure parameters on PC  Image management software (for PC) iRYS (compliant with ISDP®10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or Compliant of Processor intellection of the Intellection of	Dose delivered	Viewing on a handheld device with option of digital archive on PC via iRYS software which can be automated via the "RX DC connect" (optional) accessory
Acquisition software (for PC)  Image management software (for PC)	PC connection cable	Serial with USB adapter available in various lengths
Image management software (for PC)  iRYS (compliant with ISDP©10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 201900310 iPad iRYS viewer App (free)  Protocols supported in iRYS  ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity  iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems  Microsoft® Windows® 10. 11 Professional 64 bit  Processor  Intel Core i3 or higher  Hard Disk  100 GB SSD (250 GB recommended)  RAM  4 GB (8 GB recommended)  Discrete 3D Video Card or integrated GPU	SOFTWARE	
iPad iRYS viewer App (free)  Protocols supported in iRYS  ICOM 3.0, TWAIN, VDDS  DICOM Node Connectivity  iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log  iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems  Microsoft® Windows® 10. 11 Professional 64 bit  Processor  Intel Core i3 or higher  Hard Disk  100 GB SSD (250 GB recommended)  RAM  4 GB (8 GB recommended)  Graphics card  Discrete 3D Video Card or integrated GPU	Acquisition software (for PC)	iCapture for automatic saving of RX DC exposure parameters on PC
DICOM Node Connectivity iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)  X-ray log iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C  MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft® Windows® 10. 11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  RAM 4 GB (8 GB recommended)  Graphics card Discrete 3D Video Card or integrated GPU	Image management software (for PC)	iRYS (compliant with ISDP©10003:2020 in accordance with EN ISO/IEC17065:2012 - certificate number 2019003109-2) and iPad iRYS viewer App (free)
X-ray log iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or C MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft® Windows® 10.11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  RAM 4 GB (8 GB recommended)  Graphics card Discrete 3D Video Card or integrated GPU	Protocols supported in iRYS	ICOM 3.0, TWAIN, VDDS
MINIMUM SYSTEM REQUISITES  Supported operating systems Microsoft® Windows® 10. 11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  RAM 4 GB (8 GB recommended)  Graphics card Discrete 3D Video Card or integrated GPU	DICOM Node Connectivity	iRYS - IHE compliant (Print; Storage Commitment, SR document; WorkList; MPPS; Query/Retrieve)
Supported operating systems Microsoft® Windows® 10. 11 Professional 64 bit  Processor Intel Core i3 or higher  Hard Disk 100 GB SSD (250 GB recommended)  RAM 4 GB (8 GB recommended)  Graphics card Discrete 3D Video Card or integrated GPU	X-ray log	iRYS feature to associate exposure parameters with the X-ray images of each examination (exportable in PDF or CSV format)
Processor Intel Core i3 or higher Hard Disk 100 GB SSD (250 GB recommended) RAM 4 GB (8 GB recommended) Graphics card Discrete 3D Video Card or integrated GPU	MINIMUM SYSTEM REQUISITES	
Hard Disk 100 GB SSD (250 GB recommended)  RAM 4 GB (8 GB recommended)  Graphics card Discrete 3D Video Card or integrated GPU	Supported operating systems	Microsoft® Windows® 10. 11 Professional 64 bit
RAM 4 GB (8 GB recommended) Graphics card Discrete 3D Video Card or integrated GPU	Processor	Intel Core i3 or higher
Graphics card Discrete 3D Video Card or integrated GPU	Hard Disk	100 GB SSD (250 GB recommended)
·	RAM	4 GB (8 GB recommended)
·	Graphics card	Discrete 3D Video Card or integrated GPU
1 7 U	<u>'</u>	
Power supply  Use a power adapter of a power suitable for the video card in use	. ,	
Port USB 2.0 or later versions		


**(€** 0051




RX DC X-ray unit with HyperSphere technology

(\*) values depend on the country where the product is marketed.









www.my-ray.com



### **BU Medical Equipment**

**Plant** - Via Bicocca, 14/c - 40026 Imola - Bo (Italy) tel. +39 0542 653441 - fax +39 0542 653555

**Headquarters** - Cefla s.c. Via Selice Provinciale, 23/a - 40026 Imola - Bo (Italy) tel. +39 0542 653111 - fax +39 0542 653344

**Cefla Medical North America** 

6125 Harris Technology Blvd. - Charlotte, NC 28269 - Ph: 704 598 0020 - www.ceflamedicalna.com - info@cefladental.com





# 2 **RX DC,** X-ray unit with HyperSphere technology **FREE TO MOVE** Maximum freedom of movement with innovative ball joint (Patented).

## RX DC Hyper Technology.

Innovative design, revolutionary ergonomics, advanced technology.
RX DC - HyperSphere technology brings the best of DC X-ray units into your surgery.

MyRay, just right for you.



The wireless remote controller, the multi-mode option and the 28 adjustment levels (depending on sensor sensitivity) ensure full adaptability whatever your operating requirements.



A constant potential head tube (8 mA) with a tiny focal spot (0.4 mm at 30 cm) produces optimal images under all circumstances.



The RX DC unit features HyperSphere technology which, thanks to the full-swivel ball joint, can reach any position with ease.



Built from high quality materials and featuring a comprehensive array of equipment. Versatile and easy to install, this X-ray unit is reliable whatever the situation.



The wireless remote controller lets the user control the device (by communicating with the X-ray tube) while enjoying full freedom of movement. Access to exposure programmes is provided via two simple settings. The large display shows the sequential exposure monitor and the patient exposure dose; moreover, the controller has a wireless X-ray **snapshot** button. Wireless device control allows fast, easy installation: no fixed control

panels are required, thus providing greater freedom when positioning the X-ray unit.



### **MECHANICAL RELIABILITY**

The solid, light arms feature an effective, integrated self-balancing system that reduces any risk of tube head vibration during image acquisition.



### Hyper Ergonomy.

RX DC - HyperSphere technology allows attainment of any position with ease thanks to the revolutionary ball joint. Outstanding ergonomics ensures all your diagnostic needs are met effortlessly.

HyperSphere technology gives the RX DC unit full rotation capability. The tube revolves freely around the joint, allow it to reach practically any position, including the vertical.

RX DC - HyperSphere technology also features an automatic touch-sensitive device for simple, efficient locking/release of the X-ray head tube so it can be repositioned effortlessly between one exposure and the next. Ergonomic zones on the sides of the head provide a firm grip for effective positioning.

**Extensive positioning.** 

- electro-brake with touch-sensitive control
- infinite position range
- maximum versatility
- complete reliability





### INFINITE POSITIONS, INFINITE DIAGNOSTIC CAPABILITY

Diagnosis with unlimited movement thanks to the revolutionary ball joint which allows simple yet precise head repositioning and effortless attainment of even the trickiest positions.




### **INSTALLATION VERSATILITY**

The extruded aluminium arms - available in lengths of 40 cm, 60 cm and 90 cm to ensure outstanding installation versatility - are equipped with an integrated self-balancing system. Solid and light, they can be pointed in any direction and reduce any risk of tube head vibration during image acquisition.


### Hyper Performance.

In RX DC - HyperSphere technology, advanced ergonomics, technological innovation and revolutionary design merge to provide users with ultra-sharp images.




Sharp images at all times RX DC - HyperSphere technology provides your surgery with optimum X-ray quality whatever the type of sensor connected. Now even more powerful, with 70 kV and 8 mA, even more flexible and suitable for all commercially available sensors. The constant potential head tube, associated with the smallest intraoral imaging focal spot available (0.4 mm), ensures the best images whatever your diagnostic needs.

High definition diagnostic.




#### MULTI-MODE

Maximum flexibility to meet your diagnostic needs. Automatic parameter modulation ensures exposure power and time are always selected according to the patient's build and the specific region of investigation.



### **SEQUENCED EXPOSURE**

The dynamic service cycle allows uninterrupted use of the RX DC, as in the case of systematic examinations, and real-time monitoring of tube head temperature on the large wireless controller display.



### **RX DC CONNECT** (optional)

The RX DC X-ray unit can easily be connected to your PC via RX DC CONNECT.
Via the USB port, you can log the X-ray exposure

Via the USB port, you can log the X-ray exposure dose data in digital format.

With iRYS you can add the image to the patient's

record and the relative X-ray log. Monitor the dose value over time, display and export to other applications via shareable file.

### **MAXIMUM QUALITY**

With a tiny focal spot of 0.4 mm (at 30 cm), RX DC - HyperSphere technology produces sharp images under any condition. The tube head is now even more powerful as it operates at 70 kV, 8 mA. RX DC - HyperSphere technology gives your surgery the precision and quality of cutting-edge know-how.





### **MINIMUM DOSE**

The constant potential high frequency (DC) generator reduces the most harmful low energy radiation that is characteristic of analogue (AC) generators: current is adjustable (from 8 mA to 4 mA), as are exposure times. Moreover, the long cone (30 cm) with incorporated rectangular collimator reduces the exposed surface area. This maximises image quality and safeguards patient and worker health.



